

UNIVERSITY OF THE PHILIPPINES BAGUIO

Governor Pack Road, 2600 Baguio City

BID BULLETIN No. 1 **1 June 2021**

SUPPLY, DELIVERY AND INSTALLATION OF VARIOUS LABORATORY EQUIPMENT **Bid Reference No. UPB-BID-2021-5**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, in response to bidders' written queries received within the prescriptive period for filing.

LOCATION	ORIGINAL INSTRUCTIONS	AMENDMENT/ INCLUSION
Section VII: Technical Specifications Item no. 2	External Size: 1300x750x2250mm Internal Size (WxDxH): 1150x600x660m	External Size: Approximately 1300x750x2250mm (+-20%) Internal Size (WxDxH): Approximately 1150x600x660m (+-20%)
Section VII: Technical Specifications Item no. 3	Mini Gel Electrophoresis System	Mini Gel Electrophoresis System with compatible power supply
Section VII: Technical Specifications Item no. 8	Thermal Cycler	Gradient Thermal Cycler
	Max. Heating Rate: 6C/sec Max. Cooling Rate: 5.5C/sec	Max. Heating Rate:6C/sec to 4-6C/sec Max. Cooling Rate: 5.5C/sec to 3-5.5C/sec
Section VII: Technical Specifications Item no. 9	Dimensions: (WxDxH): at least 5.5 x 6.3 x 5.1 inc/14 x 16 x 13 cm	Dimensions: (WxDxH): approximately 5.5 x 6.3 x 5.1 inc/14 x 16 x 13 cm (+-30%)
	Weight: At least 4.84lb/2.2kg	Weight: At least 4.84lb/2.2kg

By Rules, all other related provisions in the Bidding Documents correspondingly affected by these amendments are likewise deemed amended to conform to this Bid Bulletin. Amendments made herein shall be considered an integral part of the Bidding Document.


Dr. SANTOS JOSE O. DACANAY III
Chairperson, BAC 

UNIVERSITY OF THE PHILIPPINES BAGUIO

Governor Pack Road, 2600 Baguio City

BID BULLETIN No. 1

1 June 2021

SUPPLY, DELIVERY AND INSTALLATION OF VARIOUS LABORATORY EQUIPMENT

Bid Reference No. UPB-BID-2021-5

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, in response to bidders' written queries received within the prescriptive period for filing.

LOCATION	ORIGINAL INSTRUCTIONS	AMENDMENT/ INCLUSION
Section VII: Technical Specifications Item no. 2	External Size: 1300x750x2250mm Internal Size (WxDxH): 1150x600x660m	External Size: Approximately 1300x750x2250mm (+-20%) Internal Size (WxDxH): Approximately 1150x600x660m (+-20%)
Section VII: Technical Specifications Item no. 3	Mini Gel Electrophoresis System	Mini Gel Electrophoresis System with compatible power supply
Section VII: Technical Specifications Item no. 8	Thermal Cycler	Gradient Thermal Cycler
	Max. Heating Rate: 6C/sec Max. Cooling Rate: 5.5C/sec	Max. Heating Rate: 6C/sec to 4-6C/sec Max. Cooling Rate: 5.5C/sec to 3-5.5C/sec
Section VII: Technical Specifications Item no. 9	Dimensions: (WxDxH): at least 5.5 x 6.3 x 5.1 inc/14 x 16 x 13 cm	Dimensions: (WxDxH): approximately 5.5 x 6.3 x 5.1 inc/14 x 16 x 13 cm (+-30%)
	Weight: At least 4.84lb/2.2kg	Weight: At least 4.84lb/2.2kg

By Rules, all other related provisions in the Bidding Documents correspondingly affected by these amendments are likewise deemed amended to conform to this Bid Bulletin. Amendments made herein shall be considered an integral part of the Bidding Document.


Dr. SANTOS JOSE O. DACANAY III
Chairperson, BAC